



Supplementary Figure S3: Comparison of wt% determined from x-ray intensity maps using calibration curves, to that determined by EPMA quantitative point analyses with full PAP (ZAF) correction, for a suite of known standards. A 1:1 line is included for reference. All points are close to this line, indicating minimal differences between the two methods of concentration determination.